


VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument


State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program


DataMaster cdm S/N: <u>130298</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name: <u>JAMES C. GIPSON</u>	ID#: <u>4842</u> Date: <u>21 August 09</u>
A	
Agency: <u>Palmer Police Department</u> Phone #: <u>(907) 745-4811</u>	
Instrument Location: <u>Palmer Police Department</u>	
B	
Alco S/N: <u>X172946</u>	Target Value: <u>.077</u> High Pressure: <u>1225</u>
Alco Test Value Average: <u>.075</u> <u>.073</u>	
1 st Alco 2 nd Alco	
Signature: 	
(OVER)	


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of Sept, 2009.


Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130298

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

AUGUST 21, 2009 ✓

OPERATOR'S NAME: ✓
GIPSON/JAMES/C
OPERATOR'S NUMBER: 4842
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI: ✓
RJCG8
O.L. #: 0123456
DEPT/AGENCY: PLM1
CASE/REPORT: 01-2345
TEST TYPE: U ✓
ALCO TARGET VALUE: .077 ✓
ALCO S/N: X172946 ✓

— BREATH ANALYSIS —

.077 ADJUSTED FOR 29.81 in ✓
ALCO TARGET .076 16:05
BLANK TEST .000 16:06
INTERNAL STANDARD VERIFIED 16:06
ALCO TV 29.81 in .075 ✓ 16:07
BLANK TEST .000 16:07
SUBJECT SAMPLE .000 16:08
BLANK TEST .000 16:09
ALCO TV 29.83 in .073 ✓ 16:09
BLANK TEST .000 16:10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

AUGUST 21, 2009 ✓
TIME 15:58

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMN0PQRSTUWXYZ\]^_`abcdefgijklmno
pqrstuvwxyz{|}~■